

**Corrigendum for Tender number RCB/ATN/14/15-16/DAILAB ( Multi-mode plate reader with imaging system)**

Tender Number : RCB/ATN/14 /15-16/ DAILAB

**Following changes have been made**

---

1. The specification of HTRF has been relaxed to include both Cisbio certified or Lance technology based
  
2. Point number 12 "Low Volume Plate for analysis of samples (DNA, RNA & Proteins) as low as 2ul for atleast 16 samples or more simultaneously should be made of quartz or any other low UV absorbing material"
  
3. Point number 13 - Fluorescence polarization – "System should be capable of detecting FAM label" has been included
  
4. Point number 15  
  
The absorbance wavelength is changed to – 230-1000 nm or better  
Spectral bandwidth changed to – 4 nm or better
  
5. Point number 17 - The alpha kits have been moved to optional items
  
6. Western blot detection system added to optional items
  
7. Clarification on optional items - Companies need not quote for all optional items can quote "if available"
  
8. Clarification on post warranty – "Post warranty CMC should be from the manufacturer's factory and not from the local dealer"

All other specifications and the terms and conditions of the tender remain unchanged.