

REGIONAL CENTRE FOR BIOTECHNOLOGY

180, Udyog Vihar Phase-I, Gurgaon, India

Dated 15.03.2013

CORRIGENDUM NOTICE

Please refer to the NOTICE INVITING TEDNER for Tender file No. **RCB/GTN/05/12-13 (ATPC) for the purchase Of MALDI TOF-TOF-System with Nano LC and MALDI-Spotter for Proteomic Research facility,** the undersigned is directed to incorporate the following amendment in the specification details are as under:

Particulars	EXISTING	AMENDED AS
TENDER SPECIFICATION No. C (2) (under the caption "LC- MALDI SPOTTER")	Spotting frequency should be ≤ 5 sec with user adjustable 1 second steps.	Spotting frequency should be ≤ 7 sec with user adjustable 1 second steps.
TENDER SPECIFICATION No.E (under the caption "SYSTEM CONTROL AND ANALYSIS SOFTWARE")	Comprehensive upgradable software suite for all kinds of proteomics and biomolecule applications with 2 user licences each	Comprehensive upgradable software suite for all kinds of proteomics and biomolecule applications <u>with one user licence for Data Acquisition and 2 user licence for Data Processing</u>

All the other specifications and terms & conditions of the aforementioned tender remains unaltered.

(SR. MANAGER)

For & on behalf of EXECUTIVE DIRECTOR