

Corrigendum

Tender No. RCB/ATN/11/15-16/DAILAB (Trinocular Stereozoom Microscope with Diascopic and Episcopic Illumination and Color Camera).

The following specifications have been modified:

1. Specification no. 2: The entire optics should be apochromatically corrected, with a total magnification of at least 10x-80x (with 1.0x)
2. Specification no. 9: A 5 mega pixel color high definition camera with 2/3 inch high density CCD/CMOS chip should be provided for image capture. The camera should have live display mode, USB connectivity to PC, and high frames per second rate (at least 24 frames per second at 320 x 240 resolution or better).

All other specifications and the terms and conditions of the tender remain unchanged.